

PRODUCT CONFORMITY CERTIFICATE

This is to certify that the

Smart Heated PM₁₀ BAM-1020

Manufactured by:

Met One Instruments, Inc

1600 NW Washington Blvd
Grants Pass
Oregon 97526
USA

has been assessed by CSA Group
and for the conditions stated on this certificate complies with:

**MCERTS Performance Standards for Continuous Ambient Air
Quality Monitoring Systems, Version 10 dated June 2016**

**MCERTS for UK Particulate Matter as set out in the Annex to the MCERTS Performance
Standards for Ambient Air Quality Monitoring Systems: Requirements of the UK Competent
Authority for the Equivalence Testing and Certification of Automated Continuous and
Manual Discontinuous Methods that Monitor Particulate Matter in Ambient Air, Version 1.1,
dated 31 July 2012**

Certification ranges:

PM₁₀ 0 to 1000 µg/m³

Project No.: 80006204
Certificate No: CSA MC140254/03
Initial Certification: 01 August 2014
This Certificate issued: 8 December 2022
Renewal Date: 31 July 2024



Andrew Young
Environmental Team Manager

MCERTS is operated on behalf of the Environment Agency by

CSA Group Testing UK Ltd

Unit 6, Hawarden Industrial Park
Hawarden, Deeside, CH5 3US
Tel: +44 (0)1244 670 900



0011

The MCERTS certificate consists of this document in its entirety.

For conditions of use, please consider all the information within.

This certificate may only be reproduced in its entirety and without change

To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts

Certificate Contents

Approved Site Application..... 2
 Basis of Certification 2
 Product Certified..... 3
 Certified Performance 5
 Description..... 9
 General Notes 10

Approved Site Application

Any potential user should ensure, in consultation with the manufacturer, that the monitoring system is suitable for the intended application. For general guidance on monitoring techniques refer to the Environment Agency Monitoring Technical Guidance Notes available at www.mcerts.net

The PM₁₀ field test was conducted at one site in the UK, three sites in Germany, two in Austria and one in the Czech Republic. The particulate loading at the test sites is representative of different types of areas including urban background, roadside, rural and industrial. The testing took place in both winter and summer months.

On the basis of these tests this certificate is valid when the instrument is used for urban air quality monitoring and similar applications.

The requirements according to the [Guide To The Demonstration Of Equivalence Of Ambient Air Monitoring Methods](#) (GDE 2010) are fulfilled for PM_{2.5}.

Basis of Certification

This certification is based on the following test report(s) and on CSA Group's assessment and ongoing surveillance of the product and the manufacturing process:

MCERTS certification committee Report	Certification Report and Checklist on the Evaluation of the Ambient Air Particulate Matter Monitor Test Reports for the Smart Heated PM₁₀ BAM-1020 Submitted for Approval and Certification within the MCERTS Scheme for UK Particulate Matter: Requirements of the UK Competent Authority for the Equivalence Testing of Methods that Monitor Particulate Matter in Ambient Air, 30th May 2014
Bureau Veritas	UK Report on the Equivalence of the Smart Heated PM₁₀ BAM-1020. Report ref AGGX5590185/BV/DH/2882 dated 31 May 2014
TÜV Rheinland	Report on the suitability test of the ambient air quality measuring system BAM-1020 with PM ₁₀ pre-separator of the company Met One Instruments, Inc. for the measured component suspended particulate matter PM ₁₀ TÜV-Report: 936/21205333/A dated December 06, 2006 Report published on www.qal1.de/en/hersteller/metone.htm

Certificate No: CSA MC140254/03
 This Certificate issued: 08 December 2022

*This certificate may only be reproduced in its entirety and without change
 To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

TÜV Rheinland

Addendum to the type approval test report of the measuring system BAM-1020 with PM₁₀ pre-separator of the company Met One Instruments, Inc. for the component PM₁₀ to the TÜV-report 936/21205333/A of Dec 06, 2006.

TÜV-Report: 936/21220762/A dated December 12, 2012

Report published on www.gal1.de/en/hersteller/metone.htm

Product Certified

The instrument tested was the BAM-1020 measuring system consisting of the following parts:

- USEPA style PM₁₀ sampling inlet operating at 16.67 l/min with louvered slats to prevent rain ingress
- Sampling tube
- Smart Heated Inlet (Part number BX830) set to limit the maximum relative humidity RH at the filter tape to 45 % Smart heated inlet control by Delta T was set to 'NO'
- Ambient temperature sensor (BX-592), or optionally the combined pressure and temperature sensor (BX-596)
- BAM-1020 Beta Attenuation Mass Measuring system incorporating glass fibre filter tape programmed to perform hourly measurements with a 4 minute beta attenuation measurement at the beginning and end of every 50 minute sampling period
- Vacuum pump and airflow controller set to ambient conditions

The operation of instruments in permutations other than the above is not covered by this report, and is not recommended for approval without further consideration by the UK Certification Committee for the implications of any variations.

The standard filter tape used in the Met One BAM 1020 series of particulate monitors is made from glass fibre, which is commonly available. However, the filter tape used in the BAM 1020 series of instruments has critical and customised characteristics, and as such, all filter tapes should be obtained directly from Met One or from one of its authorised dealers. Met One has historically supplied a standard filter tape (product number 460130, manufactured by Sibata in Japan). This was employed in all the seven field test programmes discussed in this certificate. Thus in principle, this type of filter tape should be employed in any future use of instruments as equivalent methods. However, Met One has more recently qualified one additional tape manufacturer (Whatman, Germany, product number 460180) which has proven to be fully equivalent in terms of its physical properties, and also in terms of the filter tape producing equivalent results for ambient atmospheric particulate monitoring of PM₁₀. The evidence supporting this has been reviewed in this Evaluation and considered to be satisfactory. This certificate requires the use of one of the designated filter tapes discussed above in order to be approved within the scope of the MCERTS for UK Particulate Matter document.

This certificate applies to all Smart Heated PM₁₀ BAM-1020 instruments with serial number D0001 onwards and fitted with Software Version 3236-02 3.2.1b onwards. It is however recommended that operators install the latest approved versions of the Software approved for the instrument; namely: Software Version 3236 V5.2.0 for those equipped with a touch screen display and Software Version 3236-07 5.1.1 for those that do not have a touch screen display. Future versions could be accepted if they are approved comprehensively during the annual audit of the manufacturing facility.

Certificate No: CSA MC140254/03
This Certificate issued: 08 December 2022

*This certificate may only be reproduced in its entirety and without change
To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Amendments to product certified

MC140254/03, dated 8 December 2022

Hardware - the addition of:

- Digital temperature sensor BX-598
- Combined pressure/temperature/humidity sensor BX-597A
- Measuring device with touch sensitive screen (BX-970, optional) with front panel USB port

Software

- The latest software version of the Smart Heated PM₁₀ BAM-1020 measuring system for PM₁₀ manufactured by Met One Instruments, Inc is: R9.3.0

Certificate No: CSA MC140254/03
This Certificate issued: 08 December 2022

*This certificate may only be reproduced in its entirety and without change
To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Certified Performance

The instrument was evaluated for use under the following conditions:

Ambient Temperature Range: 5°C to 40°C

Test	Results	MCERTS specification
Constancy of the sample volumetric flow	2.5%	To remain constant within $\pm 3\%$ of the rated value
Tightness of the sampling system	0.6% maximum	Leakage not to exceed 1% of the sampled volume
Maintenance interval	One Month	\geq Two weeks
Data Availability	98.8%	$\geq 90\%$
Number of UK Tests	1	≥ 1
Number of Reference Methods	2 for all sites but Tusimice where there was 1	≥ 1
Between sampler/instrument uncertainty for the standard method PM₁₀		
Full data set	0.59 $\mu\text{g}/\text{m}^3$	$\leq 2 \mu\text{g}/\text{m}^3$
$<30 \mu\text{g}/\text{m}^3$	0.43 $\mu\text{g}/\text{m}^3$	Not specified
$\geq 30 \mu\text{g}/\text{m}^3$	0.84 $\mu\text{g}/\text{m}^3$	Not specified

Certificate No: CSA MC140254/03
 This Certificate issued: 08 December 2022

*This certificate may only be reproduced in its entirety and without change
 To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Test	Uncorrected	Slope Corrected	MCERTS specification
Between sampler/instrument uncertainty for the candidate method PM₁₀			
Full data set	1.29 µg/m ³	1.24 µg/m ³	≤2.5 µg/m ³
<30 µg/m ³	1.14 µg/m ³	1.10 µg/m ³	≤2.5 µg/m ³
≥30 µg/m ³	1.58 µg/m ³	1.53 µg/m ³	≤2.5 µg/m ³
Expanded uncertainty calculated at 50 µg/m ³ for Instrument Austria1			
Full data set	16.0%	15.8%	≤25%
<30 µg/m ³	42.8%	33.9%	Not specified
≥30 µg/m ³	19.6%	20.1%	≤25%
Individual sites			
Graz	20.9%	20.0%	≤25%
Steyregg	9.3%	9.8%	≤25%
Expanded uncertainty calculated at 50 µg/m ³ for Instrument Austria2			
Full data set	19.8%	17.3%	≤25%
<30 µg/m ³	62.9%	52.6%	Not specified
≥30 µg/m ³	22.7%	20.7%	≤25%
Individual sites			
Graz	26.1%	22.3%	≤25%
Steyregg	9.6%	11.6%	≤25%

Certificate No: CSA MC140254/03
This Certificate issued: 08 December 2022

*This certificate may only be reproduced in its entirety and without change
To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Test	Uncorrected	Slope Corrected	MCERTS specification
Expanded uncertainty calculated at 50 µg/m ³ for Instrument J7860			
Full data set (Tusimice)	18.5%	13.2%	≤25%
<30 µg/m ³	40.5%	32.4%	Not specified
≥30 µg/m ³	17.7%	13.0%	≤25%
Expanded uncertainty calculated at 50 µg/m ³ for Instrument J7863			
Full data set (Tusimice)	18.2%	12.7%	≤25%
<30 µg/m ³	33.7%	26.1%	Not specified
≥30 µg/m ³	18.0%	12.9%	≤25%
Expanded uncertainty calculated at 50 µg/m ³ for Instrument 17011			
Full data set (Teddington)	30.0%	22.4%	≤25%
<30 µg/m ³	11.1%	5.5%	Not specified
≥30 µg/m ³	Only 1 data point	Only 1 data point	≤25%
Expanded uncertainty calculated at 50 µg/m ³ for Instrument 17022			
Full data set (Teddington)	22.3%	15.1%	≤25%
<30 µg/m ³	13.5%	7.2%	Not specified
≥30 µg/m ³	Only 1 data point	Only 1 data point	≤25%

- Note 1: The instruments tested were operated with the default measurement range of 0 to 1000 µg/m³ for hourly measurements. The maximum 24 hour average concentration recorded during the tests was 128.7 µg/m³.
- Note 2: The CM does not fulfil the relevant Data Quality Objective of EU Directive 2008/50/EC when used without correction, therefore test results were also subjected to correction for each of intercept, slope and both slope and intercept. Expanded uncertainties for the Candidate Method were calculated for both uncorrected datasets as well as data that have been adjusted for slope and/or intercept. **Slope correction of dividing by 1.035 is required in order to make the Smart Heated PM₁₀ BAM-1020 equivalent.** It is not necessary to correct for intercept, but it is essential that thorough and frequent on-going QA/QC procedures are employed (as prescribed in prEN12341:2014 and CEN/TS16450) including to precisely quantify analyser baseline performance and ensure that the instrument specific baseline correction factor programmed in to the instrument is correctly monitored and maintained.

Certificate No: CSA MC140254/03
 This Certificate issued: 08 December 2022

*This certificate may only be reproduced in its entirety and without change
 To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

PM ₁₀ Smart BAM-1020	Calculated slope of all paired data	Calculated intercept of all paired data	Expanded uncertainty of all paired data	Range of individual expanded uncertainties
Uncorrected data	1.035	0.947	17.4%	9.3% to 30.0%
Data corrected for intercept by subtracting 0.843	1.035	0.000	15.4%	9.0% to 26.3%
Data corrected for slope by dividing by 1.034	1.000	0.924	13.8%	9.8% to 22.4%
Data corrected for slope and intercept by subtracting 0.843 and then dividing by 1.034	1.000	0.009	13.4%	11.4% to 21.6%

Note 3: In order to be consistent with the 2012 TÜV Rheinland Report the instruments tested have been referred to by the serial numbers and acronyms provided by the site operators. The true serial numbers as designated by the manufacturer and the corresponding manufacture year are summarised in the following Table.

Descriptor assigned in this report	Serial number assigned by manufacturer	Build Year
Austria 1	G8210	2007
Austria 2	G8211	2007
J7860	J7860	2009
J7863	J7863	2009
17011	G2757	2007
17022	G4044	2007

Note 4: A study of pollution climate relevant to sites in the UK, the Czech Republic, and Austria has demonstrated that in all cases the particulate geometric mean criteria are met and at least one site meets the lower threshold and higher threshold criterion for wind speed, ambient temperature, ambient dew point and semi volatile nitrate content. The pollution climate criteria are satisfied for all the equivalence tests.

Note 5: For the purposes of quality control of these monitors in the field, as with all PM instruments that are not the reference method, the Smart Heated PM₁₀ BAM-1020 should be calibrated on a test site at intervals against the gravimetric reference methods EN 12341 or EN 14907 as applicable, and as given in the recommendations of the GDE 2010 and CEN/TS16450.

Note 6: Three German datasets of Cologne Parking Lot; Titz-Rödingen; and Cologne Frankfurter Strasse each had fewer than the required 40 data points, but were included in the 2012 TÜV Rheinland Report at the request of the German Certification Committee. The conclusion of the TÜV Rheinland Report when considering all seven datasets was that slope correction of dividing by 1.034 is required in order to make the instrument equivalent. For UK purposes it is necessary to consider only those datasets with greater than 40 data points. It has been shown that when considering only the four datasets with greater than 40 data points that it is required to slope correct by a slightly different constant; namely: **slope correction of dividing by 1.035 is required in order to make the Smart Heated PM₁₀ BAM-1020 equivalent.**

Note 7: A comprehensive set of laboratory tests were undertaken by TÜV Rheinland and are discussed in the TÜV Rheinland report. It is not necessary to review the results of these tests under the MCERTS for UK Particulate Matter certification scheme.

Certificate No: CSA MC140254/03
 This Certificate issued: 08 December 2022

*This certificate may only be reproduced in its entirety and without change
 To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Description

The Smart Heated PM₁₀ BAM-1020 Ambient Particulate Monitor is used to sample PM₁₀ by using a 16.7 l/min USEPA style PM₁₀ inlet. The system uses smart heating technology in order to limit the maximum relative humidity RH at the filter tape to 45 %. The instrument utilizes glass fibre tape.

The ambient air measuring system BAM-1020 is based on the measuring principle of beta-attenuation. The principle of the radiometric determination of mass is based on the physical law of attenuation of beta-rays when passing a thin layer of material. There is the following relationship:

$$C = \frac{10^6 A (\text{cm}^2) \ln \frac{I_0}{I}}{Q C_e D t (\text{min}) m_g} \text{ g m}^{-3}$$

where:

- C particle-mass concentration;
- A sampling area for particles (filter spot);
- Q sampling flow rate;
- Dt sampling time;
- μ mass absorption coefficient;
- I₀ beta count rate at the beginning (clean);
- I beta count at the end (collect).

The radiometric determination of mass is calibrated in the factory. During routine operation of the instrument this is checked hourly both on the clean filter prior to collection of the sample and using the built-in reference foil. The values obtained can be compared with any stability requirements (such as drift effects) relative to the values obtained during factory calibration.

One measurement cycle (incl. automatic check of the radiometric measurement) consists of the following steps (setting: measuring time for radiometry 4 min):

1. The initial count of the clean filter tape I₀ is performed at the beginning of the cycle for a period of four minutes.
2. The filter tape is advanced four windows and the sampling (vacuum pumping) begins on the spot in which I₀ was just measured. Air is drawn through this spot on the filter tape for approximately 50 minutes.
3. At the same time the second count I₁ occurs (at a point on the tape 4 windows back) for a period of four minutes. The purpose of the measurement is to perform the verification for instrument drift caused by varying external parameters such as temperature and relative humidity. A third count I₂ occurs with the reference membrane extended over the same place on the tape. Four minutes before the end of sampling time, another count I_{1x} occurs on the same point of the tape. With the help of I₁ and I_{1x}, the stability at the zero point can be monitored.

Certificate No: CSA MC140254/03
 This Certificate issued: 08 December 2022

*This certificate may only be reproduced in its entirety and without change
 To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

4. After sampling, the filter tape is moved back four windows to measure the beta ray absorption through the section that has collected dust (I3). Finally the concentration calculation is performed to complete the cycle.
5. The next cycle begins with step 1.

General Notes

1. This certificate is based upon the equipment tested. The Manufacturer is responsible for ensuring that on-going production complies with the standard(s) and performance criteria defined in this certificate. The manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management system shall be subject to regular surveillance according to 'Regulations Applicable to the Holders of CSA Group Testing UK Ltd Certificates'.
2. The design of the product certified is held and maintained by TÜV Rheinland for certificate No. Sira MC140254/03.
3. If a certified product is found not to comply, CSA Group should be notified immediately at the address shown on this certificate.
4. The certification marks that can be applied to the product or used in publicity material are defined in 'Regulations Applicable to the Holders of CSA Group Testing UK Ltd Certificates'.
5. This document remains the property of CSA Group and shall be returned when requested by CSA Group.

Certificate No: CSA MC140254/03
This Certificate issued: 08 December 2022

*This certificate may only be reproduced in its entirety and without change
To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*