

# PRODUCT CONFORMITY CERTIFICATE

This is to certify that the

**WATERFLUX 3100**

Manufactured by:

**KROHNE Altometer**

*A production facility of KROHNE AG, Basel*

*Kerkeplaat 12*

*3313 LC Dordrecht*

*The Netherlands*

has been assessed by Sira Certification Service  
and for the conditions stated on this certificate complies with:

**MCERTS Performance Standards for Water Monitoring  
Equipment Part 3, Version 2.4 February 2013**

The combined performance characteristic ( $U_c$ , the expanded uncertainty) is **1.57%** (Class1)

Certification Ranges:

Size range DN25 to DN500

Project No.: 16W22202/ 70157923  
Certificate No: Sira MC120207/03  
Initial Certification: 25 September 2012  
This Certificate issued: 17 March 2021  
Renewal Date: 24 September 2022



Andrew Young  
Environmental Project Engineer

MCERTS is operated on behalf of the Environment Agency by

**Sira Certification Service**

Unit 6, Hawarden Industrial Park  
Hawarden, Deeside, CH5 3US  
Tel: +44 (0)1244 670 900



0011

**sira**  
CERTIFICATION

*The MCERTS certificate consists of this document in its entirety.  
For conditions of use, please consider all the information within.  
This certificate may only be reproduced in its entirety and without change  
To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*

## Approved Site Application

The product may be used on all MCERTS applications including abstraction, effluent discharge, ultraviolet disinfection and industrial processing.

*Any potential user should ensure, in consultation with the manufacturer, that the monitoring system is suitable for the intended application. For general guidance on monitoring techniques refer to the Environment Agency Monitoring Technical Guidance Notes available at [www.mcerts.net](http://www.mcerts.net).*

## Basis of Certification

This certification is based on the following Test Report(s) and on Sira's assessment and ongoing surveillance of the product and the manufacturing process:

Sira Report	16W22202 V03 dated 11 September 2012
NMi Certain B.V Report	R49-1/2003-NL1 – 06.01 dated 17 March 2006

## Product Certified

The WATERFLUX 3100 consists of the following parts:

- WATERFLUX 3000 electromagnetic flow / water meter
- IFC 100 electromagnetic signal converter (C / compact or F / remote version)

This certificate applies to all instruments fitted with software version 4.0.4. (serial number A10 01 xxxxx onwards).

Pipe size	Flow rate		Unit
	max	min	
DN25	16	0.32	m <sup>3</sup> /hr
DN40	25	0.50	m <sup>3</sup> /hr
DN50	40	0.81	m <sup>3</sup> /hr
DN65	63	1.30	m <sup>3</sup> /hr
DN80	100	2.00	m <sup>3</sup> /hr
DN100	160	3.20	m <sup>3</sup> /hr
DN125	250	5.00	m <sup>3</sup> /hr
DN150	400	8.10	m <sup>3</sup> /hr
DN200	630	13.00	m <sup>3</sup> /hr
DN250	1000	20.00	m <sup>3</sup> /hr
DN300	1600	32.00	m <sup>3</sup> /hr
DN350	2500	12.70	m <sup>3</sup> /hr
DN400	4000	20.32	m <sup>3</sup> /hr
DN450	4000	20.32	m <sup>3</sup> /hr
DN500	6300	32.00	m <sup>3</sup> /hr

Certificate No:	Sira MC120207/03
This Certificate issued:	17 March 2021

*This certificate may only be reproduced in its entirety and without change  
To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*

## Certified Performance

The instrument was evaluated for use under the following conditions:  
Ambient Temperature Range: -25°C to +55°C

The instrument meets MCERTS Class 1 requirements for the combined performance characteristic as specified in Table 6 of the MCERTS performance standard. Details of individual performance characteristics are summarised below:

Results are expressed as error % reading, unless otherwise stated.

Test	Results expressed as error % of reading				Other results	MCERTS specification
	<0.5	<1.0	<1.5	<2.0		
Protection against unauthorised access	A custody transfer seal is present					Clause 3.1.2
Indicating device	The flowmeter incorporates an indicating device, analogue and digital output signal					Clause 3.1.3
Units of measurement	Various units of measurement are available and displayed.					Clause 3.1.6
Bi-directional flow	The flowmeter displays a '+' or '-' flow reading					Clause 3.1.8
<b>Combined performance characteristic</b>				<b>1.57</b>		Clause 6.3.2 ±2.0% <b>Class 1</b>
Mean error						
DN25	0.34					
DN80		-0.99				
DN250	-0.25					
DN400			1.07			Clause 6.3.2 ±1.5% Class 1
DN500	-0.24					
Repeatability						
DN25	0.31					
DN80	0.36					
DN250	0.24					
DN400	0.44					
DN500	0.14					
Supply voltage	<0.02				70 to 250 VAC 10 to 26.4 VDC	Clause 6.3.3 0.5% Class 1
Output impedance	<0.02				50 to 1000Ω	Clause 6.3.4 0.5% Class 1
Fluid Temperature	-0.33				12°C to 50°C	Clause 6.3.5 0.5% Class 1
Ambient air temperature	0.01				-25°C to +55°C	Clause 6.3.6 0.5% Class 1
Relative humidity	0.01				Test conducted at 45% relative humidity	Clause 6.3.6 0.5% Class 1
Stray currents	<0.02					Clause 6.3.9 0.5% Class 1

Certificate No: Sira MC120207/03  
This Certificate issued: 17 March 2021

*This certificate may only be reproduced in its entirety and without change  
To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*

Test	Results expressed as error % of reading				Other results	MCERTS specification
	<0.5	<1.0	<1.5	<2.0		
Bi-directional flow Mean error Repeatability	-0.38				See Note 1	Mean error ±1.5% Class 1 Repeatability 1% Class 1
Loss of Power for electronic flowmeters	No changes in pre set data					Clause 6.3.1 to be reported
Response time					<10s	Clause 6.3.19 30 seconds
Warm up time					<1s	Clause 6.1.2 to be reported
Vibration					Note 2	Clause 6.3.20 to be reported

Note 1: Repeatability for bi-directional flow could not be calculated as only 2 data points were taken at each flow rate

Note 2: Test not conducted

Note 3: The following tests are not applicable to the flowmeter:

6.3.4	Output Impedance	6.3.14	Flow reversal
6.3.7	Incident light	6.3.15	Ancillary devices
6.3.8	Sensor location	6.3.16	Effect of conduit material
6.3.10	Sonic velocity compensation & response	6.3.17	Effect of conduit size
6.3.11	Accuracy of computation	6.3.18	Fill level
6.3.12	User defined stage-discharge equation		

Certificate No: Sira MC120207/03  
This Certificate issued: 17 March 2021

*This certificate may only be reproduced in its entirety and without change  
To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*

## Field Test Results

The field test was conducted on a WATERFLUX 3100 in series with an electromagnetic flowmeter for 3 months measuring leachate at a landfill site

Test	Results expressed as error % of reading				Other results	MCERTS specification
	<0.5	<1.0	<1.5	<2.0		
Error under field test conditions	Error range -1.54% to 0.00%					Clause 7.3  2% Class 1 5% Class 2 8% Class 3
Up time					100%	Clause 7.4 >95%
Maintenance					none	Clause 7.5 to be reported

Certificate No: Sira MC120207/03  
This Certificate issued: 17 March 2021

*This certificate may only be reproduced in its entirety and without change  
To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*

## Description:

The WATERFLUX 3100 is a mains powered water meter based on the electromagnetic flow principle. It is dedicated for applications in the water and wastewater industry. It consists of a WATERFLUX 3000 sensor and a mains powered IFC 100 converter. In terms of available housing versions, there is a compact variant, in which the signal converter is connected to the measuring sensor, as well as a 0° and 45° version. The converter can be mounted directly on the sensor (compact version) or separately (field version).

WATERFLUX 3100 it is certified with DVGW, ACS, KTW for potable water applications. Krohne Altometer meets the applicable requirements of MID module D for the conformity assessment of water meters. Every flow meter leaving the factory is calibrated.

The principle of the electromagnetic flow metering is based on Faraday's law of induction: passing an electrically conductive body through a magnetic field, a voltage is induced. This voltage is proportional to velocity and picked up by electrodes.

The IFC 100 has been equipped with a wide variety of diagnostic tools for device function and application check. Conductivity measurement, Electrode error & Process or ambient temperature too high

The WATERFLUX 3000 sensor has a rectangular cross section for sizes DN25 to DN300 resulting in an improved flow profile, high accuracy in low flow conditions, large span and minimal power consumption. It is stated by the manufacturer that straight up and downstream lengths are not required. Above DN300 the Waterflux 3000 sensor is full bore.

The IFC 100 converter is mains powered. It can display positive and negative counter, sum counter and flow rate. It can further provide diagnostic information for self checking, counter overrun, flow direction and empty pipe detection. Available inputs and outputs: Current output (incl. HART®), pulse/frequency output, status output and control input.

## General Notes

1. This certificate is based upon the equipment tested. The Manufacturer is responsible for ensuring that on-going production complies with the standard(s) and performance criteria defined in this Certificate. The Manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management system shall be subject to regular surveillance according to 'Regulations Applicable to the Holders of Sira Certificates'. The design of the product certified is defined in the Sira Design Schedule for certificate No. Sira MC120207/02
2. If certified product is found not to comply, Sira Certification Service should be notified immediately at the address shown on this certificate.
3. The Certification Marks that can be applied to the product or used in publicity material are defined in 'Regulations Applicable to the Holders of Sira Certificates'.
4. This document remains the property of Sira and shall be returned when requested by the company.

Certificate No: Sira MC120207/03  
This Certificate issued: 17 March 2021

*This certificate may only be reproduced in its entirety and without change  
To authenticate the validity of this certificate please visit [www.csagroupuk.org/mcerts](http://www.csagroupuk.org/mcerts)*